Test Equipment Solutions Datasheet

Test Equipment Solutions Ltd specialise in the second user sale, rental and distribution of quality test & measurement (T&M) equipment. We stock all major equipment types such as spectrum analyzers, signal generators, oscilloscopes, power meters, logic analysers etc from all the major suppliers such as Agilent, Tektronix, Anritsu and Rohde & Schwarz.

We are focused at the professional end of the marketplace, primarily working with customers for whom high performance, quality and service are key, whilst realising the cost savings that second user equipment offers. As such, we fully test & refurbish equipment in our in-house, traceable Lab. Items are supplied with manuals, accessories and typically a full no-quibble 1 year warranty. Our staff have extensive backgrounds in T&M, totalling over 150 years of combined experience, which enables us to deliver industry-leading service and support. We endeavour to be customer focused in every way right down to the detail, such as offering free delivery on sales, presenting flexible technical + commercial solutions and supplying a loan unit during warranty repair, if available.

As well as the headline benefit of cost saving, second user offers shorter lead times, higher reliability and multivendor solutions. Rental, of course, is ideal for shorter term needs and offers fast delivery, flexibility, try-before-you-buy, zero capital expenditure, lower risk and off balance sheet accounting. Both second user and rental improve the key business measure of Return On Capital Employed.

We are based at Aldermaston in the UK from where we supply test equipment worldwide. Our facility incorporates Sales, Support, Admin, Logistics and our own in-house Lab.

All products supplied by Test Equipment Solutions include:

- No-quibble parts & labour warranty (we provide transport for UK mainland addresses).
- Free loan equipment during warranty repair, if available.
- Full electrical, mechanical and safety refurbishment in our 40GHz in-house Lab.
- Certificate of Conformance (calibration available on request).
- Manuals and accessories required for normal operation.
- Free insured delivery to your UK mainland address (sales).
- Support from our team of seasoned Test & Measurement engineers.
- ISO9001 quality assurance.

Test Equipment Solutions Ltd Unit 3 Zodiac House Calleva Park Aldermaston Berkshire RG7 8HN

T: 01183 800 800 F: 01183 800 804

Email: info@TestEquipmentHQ.com
Web: www.TestEquipmentHQ.com






PS-19/PSS-19 **Level Generator**

1. July 1

for the frequency range 80 Hz to 25 MHz

IEEE 488 IEC 625)



- Crystal controlled frequency settings through keyboard; stepwise or continuous
- High accuracy of frequency, max. resolution 0.1 Hz
- Send level adjustable in 0.1 dB steps
- Level displayed in dB/dB0 or in dBm/dBm0
- Coaxial (75 Ω or 50 Ω) and balanced outputs
- Slave tuning by SPM-19 Level Meter (SPM-18)
- Sweepable send frequency, if requested
- Storage and call-up of any fixed frequencies, and a menu of complete instrument settings
- Economical Measuring Setup design with PSS-19 Send Section and SPM-19 Level Meter (SPM-18)
- Point-by-point and sweep measurements with frequency offset (PS-19/SPM-19)
- Extendable for use as level standard

Applications

The PS-19 Level Generator provides a signal source for use in the development, manufacture, installation and maintenance of balanced and coaxial FDM systems with up to 3600 voice-grade channels. It can also be used for measurements in the lower multiplexer range of single sideband radio-link systems and in individual telephone channels.

Because of its remote control facility, the PS-19 can be used in automatic test systems. When used in conjunction with the SPM-19 Level Meter a complete measuring setup for level, gain and loss measurements is created.

The setup can be provided with a sweep facility on request. Measurements using frequency offsets, e.g. on translators, are possible. The instrument is particularly suitable for selective end-to-end measurements on VFT systems.

The PS-19's send frequency is derived from a synthesizer, which has a remarkably high spectral purity, accuracy, stability and freedom from phase-hits at frequency changeovers (this applies particularly to sweep measurements).

The send frequency on the PS-19 can be entered digitally via a keyboard, a continuous adjustment control or an increment key. The resolution is 0.1 Hz for all three methods.

To all intents and purposes, any number of fixed frequencies and PS-19 settings can be stored to facilitate repetitive measurements.

The built-in data retention capability ensures that the memory contents are not lost if there is a power failure. The send level, which may be "soft" blanked, can be set very precisely in 0.1 dB steps. After the microprocessor controlled, fully programmable Level Generator has been switched on, a self-check is carried out to ensure the PS-19 will function correctly. The PS-19 has female connectors for WECO male connectors.

Frequency range, coaxial 80 Hz to 25 MHz balanced 80 Hz to 14 MHz
Frequency setting via keyboard, quasi-continuous,
by increments manual or automatic, frequency sweep (on request)
Frequency error limits $\dots \pm 3 \times 10^{-7}$ or $\pm 1 \times 10^{-7}$ Level range,
coaxial (75Ω) -83.9 to $+1$ dB $(-74.9$ to $+10$ dBm) Send level error at 20 kHz and $Z_o=75\Omega$ ±0.1 dB
Output impedances 0, 75 *), 124, 150 (135), 600 Ω Storable fixed frequencies (100/200) and PS-19 settings (10/50)

^{*) 50 \}Omega on request


Further Characteristics and Applications

- High frequency accuracy: The very small uncertainty of 3×10^{-7} on the set frequency (1×10^{-7} as option), means that frequencies can be set very accurately even at the high frequencies required for making in-service measurements in the channel gaps of FDM systems, for example. The error stated is valid in the rated range of use for temperature and includes ageing of the reference crystal.
- Sweep capability: The send frequency can be swept over the whole of the PS-19's frequency range (BN 870/02 and /22). The rapid and easy-to-set sweep limits are particularly useful for wideband end-to-end measurements. The centre frequency and sweep width are also adjustable.
- Storage of fixed frequencies and PS-19 settings: The possibility of calling up 100 (with User-specific Memory 200) fixed frequencies and 10 or 50 PS-19 settings takes the tedium out of repetitive measurements. The fixed frequencies can be stepped through in sequence automatically or manually.
 - If the SPM-19 (or SPM-16) is connected at the remote site, selective end-to-end measurements can be carried out in this operating mode without additional synchronisation. When a frequency is changed it is possible to "soft" blank the send level.
- Measurement with frequency shift: Point-by-point and sweep measurements using frequency offsets, e.g. on translators, are possible with the PS-19/SPM-19 setup. The

send and receive frequencies can be detuned in the same or opposite direction by an amount equal to the frequency offset.

1

- Digital level display: This means that the send level can be set quickly and precisely with high resolution. The absolute level referred to a point of zero relative level (dBm0 or dB0) and the relative level of the measurement point (dBr) can also be set for measurements on FDM systems.
- Use as a level standard: The PS-19 can be used as a
 precise a.c. signal source producing a level of 0 dBm. This is
 achieved by forming a control loop in which the EPM-1 Milliwatt Power Meter acts as an ALC amplifier and feeds a
 control voltage to the control voltage input of the PS-19. This
 arrangement eliminates the loss and matching errors introduced by long leads.
- Economical design: Only one tuning oscillator is required for frequency setting when measurements with the same send and receive frequency are carried out. The PSS-19 Send Section/SPM-19 Level Meter combination is ideal for measurements of this kind.
- SNZ-1 Standard Frequency Adaptor Option is used for external synchronisation using frequencies between 0.3 and 9.9 MHz (100 kHz steps). Power supply from separate mains unit (e.g. BN 964/00.0X; 12 V/50 mA).



g die g

Unless otherwise noted, the specifications are valid for the rated range of use of the a.c. line voltage, a.c. line frequency and the ambient temperature 15 min after the set is switched on. The level ranges for dBm calibration are shown in brackets.

Outputs Output Versacon*s g digital through keyboard, resolution 0.1 dB Coaxial output* Versacon*s 9 Universal Connector System adaptable to all commercially used connectors and properties of the properties of t
Universal Connector System adaptable to all commercially used connectors $\mathbb{B} \times \mathbb{B} \times \mathbb{B}$
BN 870/21 and /22 Fem. com. for WECO 38S A male com. Output impedance 75 $_{20}$ r 50 $_{20}$ Frequency range 80 Hz to 25 MHz t
BN 870/21 and /22-Fam. com. for WEOO 38 A male sonn.
Coaxial output (so Ω on request) Frequency range
Frequency range $80 \text{Hz} \text{to} 25 \text{MHz}$ B al an c ed out put s^{11} $3 \text{pole} \text{CF}$ jacks Output impedance, switchable $124 \Omega_1 150 \Omega_2$ Frequency range $60 \text{kHz} \text{to} 14 \text{MHz}$ Signal balance ratio according to CCITT Rec. 0.121 $\geq 40 \text{dB}$ Output impedance, switchable $150 \Omega_2 600 \Omega_2 \text{approx} 0.0 \Omega_2$ Frequency range $80 \text{Hz} \text{to} 14 \text{MHz}$ Signal balance ratio $(-75.3 \text{to} +9.6 \text{dBm})$ Output impedance, switchable $150 \Omega_2 600 \Omega_2 \text{approx} 0.0 \Omega_2$ Frequency range $80 \text{Hz} \text{to} 620 \text{kHz}$ Signal balance ratio $(-75.3 \text{to} +9.6 \text{dBm})$ Dutput impedance, switchable $150 \Omega_2 600 \Omega_2 \text{approx} 0.0 \Omega_2$ Frequency range $80 \text{Hz} \text{to} 620 \text{kHz}$ Signal balance ratio $(-75.3 \text{to} +9.6 \text{dBm})$ Dutput impedance, switchable $150 \Omega_2 600 \Omega_2 \text{approx} 0.0 \Omega_2$ Frequency range $80 \text{Hz} \text{to} 620 \text{kHz}$ Signal balance ratio $(-75.3 \text{to} +9.6 \text{dBm})$ Balanced output II $(-75.2 \text{to} +2.7 $
Balanced outputs 0 3pole CF jacks Output impedance, switchable 124 Ω , 150 Ω Frequency range 60 kHz to 14 MHz Signal balance ratio according to CCITT Rec. 0. 121 \geq 40 dB Output impedance, switchable 150 Ω , 600 Ω , approx 0. Ω Frequency range 80 Hz to 620 kHz Signal balance ratio according to CCITT Rec. 0. 121 \geq 40 dB Output impedance, switchable 150 Ω , 600 Ω , approx 0. Ω Frequency range 80 Hz to 620 kHz Signal balance ratio 40 dB Signal balance ratio 51 Ω cording to CCITT Rec. 0. 121 \geq 40 dB Signal balance ratio 52 Ω cording to CCITT Rec. 0. 121 \geq 40 dB Signal balance ratio 53 Ω cording to CCITT Rec. 0. 121 \geq 40 dB Signal balance ratio 64 Ω cording to CCITT Rec. 0. 121 \geq 40 dB Signal balance ratio 65 Ω cording to CCITT Rec. 0. 121 \geq 40 dB Signal balance ratio 67 Ω cording to CCITT Rec. 0. 121 \geq 40 dB Signal balance ratio 68 Ω cording to CCITT Rec. 0. 121 \geq 40 dB Signal balance atting a PS-19, digital keyboard, resolution 0. 1 Hz 67 Ω cordinates without 5 and 5 switching manual tuning over the whole range without 5 and 5 switching smallest step 1 Hz 67 Ω cordinates with 0 ption Device BN865/00.03 Ω to 1 Hz 7 Ω cordinates Ω
Frequency range 60kHz to 14MHz Signal balance ratio according to CCITT Rec. 0.121 $\geq 40\text{dB}$ Cutput impedance, switchable 150Ω , 600Ω , approx, 0Ω Frequency range 80Hz to 620kHz bignal balance ratio according to CCITT Rec. 0.121 $\geq 40\text{dB}$ Signal balance ratio according to CCITT Rec. 0.121 $\geq 40\text{dB}$ according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $2_{\text{out}} = Z_L = 150\Omega$ $-83.9\text{to} + 11\text{dB}$ Signal balance ratio according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$) according to CCITT Rec. 0.121 $\geq 40\text{dB}$ $-78\text{to} + 17\text{dBm}$ according to CCITT Rec. 0.121 $= 40\text{dB}$ $-78\text{to} + 17\text{dBm}$ according to CCITT Rec. 0.121 $= 40\text{dB}$ $-78\text{to} + 17\text{dBm}$ according to CCITT Rec. 0.121 $= 40\text{dB}$ according to CITT. To 0.121 $= 40\text{dB}$ according to CITT. To 0.12
Signal balance ratio according to CCITT Rec. 0.121 . $\geq 40\mathrm{dB}$ according to CCITT Rec. 0.121 . $\geq 40\mathrm{dB}$ according to CCITT Rec. 0.121 . $\geq 40\mathrm{dB}$ ($-76.2\mathrm{to} + 8.7\mathrm{dBm}$) balanced output II $Z_{\mathrm{out}} = Z_L = 150\Omega$. $-83.9\mathrm{to} + 11\mathrm{dB}$ balanced output II $Z_{\mathrm{out}} = Z_L = 150\Omega$. $-83.9\mathrm{to} + 11\mathrm{dB}$ ($-77.9\mathrm{to} + 17\mathrm{dBm}$) according to CCITT Rec. 0.121 . $\geq 40\mathrm{dB}$ ($-78.2\mathrm{dB}$) balanced output II $Z_{\mathrm{out}} = Z_L = 150\Omega$. $-83.9\mathrm{to} + 11\mathrm{dB}$ ($-77.9\mathrm{to} + 17\mathrm{dBm}$) according to CCITT Rec. 0.121 . $\geq 40\mathrm{dB}$ ($-78.2\mathrm{dB}$) balanced output II $Z_{\mathrm{out}} = Z_L = 150\Omega$. $-83.9\mathrm{to} + 11\mathrm{dB}$ ($-77.9\mathrm{to} + 17\mathrm{dBm}$) according to CCITT Rec. 0.121 . $\geq 40\mathrm{dB}$ ($-78.2\mathrm{dB}$) balanced output II $Z_{\mathrm{out}} = Z_L = 150\Omega$. $-83.9\mathrm{to} + 11\mathrm{dB}$ ($-77.9\mathrm{to} + 17\mathrm{dBm}$) according to CCITT Rec. 0.121 . $\geq 40\mathrm{dB}$ ($-78.2\mathrm{dB}$) balanced output II $Z_{\mathrm{out}} = Z_L = 150\Omega$. $-78.9\mathrm{to} + 17\mathrm{dBm}$) according to CCITT Rec. 0.121 . $\geq 40\mathrm{dB}$ ($-78.2\mathrm{dB}$) balanced output II $Z_{\mathrm{out}} = Z_L = 150\Omega$. $-78.9\mathrm{to} + 17\mathrm{dBm}$) according to CCITT Rec. 0.121 . $\geq 40\mathrm{dB}$ ($-78.2\mathrm{dB}$) balanced output II $Z_{\mathrm{out}} = Z_L = 150\Omega$. $-78.9\mathrm{to} + 17\mathrm{dBm}$) according to CCITT Rec. 0.121 . $\geq 40\mathrm{dB}$ ($-78.2\mathrm{dB}$) balanced output II $Z_{\mathrm{out}} = Z_L = 150\Omega$. $-78.9\mathrm{to} + 17\mathrm{dBm}$) according to CCITT Rec. 0.121 . $-78.9\mathrm{to} + 17\mathrm{dBm}$) according to CCITT Rec. 0.121 . $-78.9\mathrm{to} + 17\mathrm{dBm}$) according to CCITT Rec. 0.121 . $-78.9\mathrm{to} + 17\mathrm{dBm}$) according to CCITT Rec. 0.121 . $-78.9\mathrm{to} + 17\mathrm{dBm}$) according to CCITT Rec. 0.121 . $-78.9\mathrm{to} + 17\mathrm{dBm}$) according to CCITT Rec. 0.121 . $-78.9\mathrm{to} + 17\mathrm{dBm}$) according to CCITT Rec. 0.121 . $-78.9\mathrm{to} + 17\mathrm{dBm}$) according to CCITT Rec. 0.121 . $-78.9\mathrm{to} + 17\mathrm{dBm}$) according to CCITT Rec. 0.121 . $-78.9\mathrm{to} + 17\mathrm{dBm}$) according to CCITT Rec. 0.121 . $-78.9\mathrm{to} + 17\mathrm{dBm}$) according to CITT Rec. 0.121 . $-78.9\mathrm{to} + 17dBm$
Output impedance, switchable 150Ω , 600Ω , approx. 0Ω Frequency range 80Hz to 620kHz Signal balance ratio according to CCITT Rec. O. 121
Dutput impedance, switchable . 150 Ω , 600 Ω , approx. 0 Ω Frequency range . 80 Hz to 620 kHz Signal balance ratio . (-77.9 to +17 dBm) according to CCITT Rec. O. 121 . \geq 40 dB . $2_{out} = Z_L = 150 \Omega$. $-83.9 \text{to} +11 \text{dB}$. $(-77.9 \text{to} +17 \text{dBm})$ according to CCITT Rec. O. 121 . \geq 40 dB . $(-77.9 \text{to} +17 \text{dBm})$. $(-77.9 \text{to} +16.9 \text{dB})$. $(-72 \text{to} +22.9 \text{dBm})$. $(-77.9 \text{to} +17 \text{dBm}$
Frequency range 80Hz to 620kHz Signal balance ratio according to CCITT Rec. O. 121 $\ge 40\text{dB}$ 20mt
Signal balance ratio according to CCITT Rec. O. 121 \geq 40 dB $> 2_{out} = 0$, $Z_L = 150 \Omega = -78$ to $+17$ dBm) according to CCITT Rec. O. 121 \geq 40 dB $> 2_{out} = 0$, $Z_L = 150 \Omega = -78$ to $+16.9$ dBm. Frequency setting at PS-19, digital keyboard, resolution 0.1 Hz quasi-continuous manual tuning over the whole range without band switching smallest step 1.1 Hz in steps by increment keys, smallest increment 1.1 Hz Error limits of frequency including ageing over 1 year 1.0^{-7} with Option Device BN 865/00.03 1.1×10^{-7} Slave tuning 1.5 through Level Meters SPM-19, SPM-18 Frequency setting at PSS-19 through 1.5 Level Meters SPM-19, SPM-18 Automatic frequency adjustments Auto-step Automatic increment, limits and speed of incrementation via keyboard. Swe ep mode with PS-19, BN 870/02 and BN 870/22 with PSS-19 through SPM-19, BN 829/02, /03 or /22 Setting of upper and lower sweep limits or setting of centre frequency and sweep width Sweep 1.5 periodic (triangular), single shot, and manual speed of intergrand seven with the position of centre frequency and sweep width Sweep 1.5 periodic (triangular), single shot, and manual speed of line of the position of the position of the position of the position of centre frequency and sweep width Sweep 1.5 periodic (triangular), single shot, and manual speed of centre frequency and sweep width Sweep 1.5 periodic (triangular), single shot, and manual speed of line of the position of t
Frequency Frequency setting at PS-19, digital keyboard, resolution 0.1 Hz quasi-continuous manual tuning over the whole range without band switching smallest step 1 Hz in steps by increment keys, smallest increment 1 Hz Error limits of frequency including ageing over 1 year $\pm 3 \times 10^{-7}$ with Option Device BN 865/00.03 $\pm 1 \times 10^{-7}$ Slave tuning through Level Meters SPM-19, SPM-18 Frequency setting at PSS-19 through Level Meters SPM-19, SPM-18 Automatic frequency adjustments Automatic incrementation of the frequency between two limits entered on the PS-19. Entry of increment, limits and speed of incrementation via keyboard. Sweep mode with PSS-19 through SPM-19, BN 870/02 and BN 870/22 with PSS-19 through SPM-19, BN 829/02, /03 or /22 Setting of upper and lower sweep limits or setting of centre frequency and sweep width Sweep periodic (triangular), single shot, and manual strong and lower seep width Sweep periodic (triangular), single shot, and manual strong and lower seep width Sweep periodic (triangular), single shot, and manual strong and lower seep limits or setting of centre frequency and sweep width Sweep periodic (triangular), single shot, and manual strong through setting of centre frequency and sweep width Sweep periodic (triangular), single shot, and manual strong setting of centre frequency and sweep width sweep setting at PS-19, single shot, and manual strong setting sett
Frequency Setting at PS-19, digital keyboard, resolution 0.1 Hz quasi-continuous manual tuning over the whole range without band switching smallest step 1 Hz in steps by increment keys, smallest increment 1 Hz Error limits of frequency including ageing over 1 year 43 × 10 − with Option Device BN 865/00.03 ±1 × 10 − Slave tuning through Level Meters SPM-19, SPM-18 Frequency setting at PSS-19 through Level Meters SPM-19, SPM-18 Auto-step Automatic incrementation of the frequency between two limits entered on the PS-19. Entry of increment, limits and speed of incrementation via keyboard. Sweep mode with PS-19, BN 870/02 and BN 870/22 with PSS-19 through SPM-19, BN 829/02, /03 or /22 Setting of upper and lower sweep limits or setting of centre frequency and sweep width Sweep — periodic (triangular), single shot, and manual state of the PS - 19. Periodic (triangular), single shot, and manual state of the PS - 19. Periodic (triangular), single shot, and manual state of the PS - 19. Periodic (triangular), single shot, and manual state of the PS - 19. Periodic (triangular), single shot, and manual state of the PS - 19. Periodic (triangular), single shot, and manual state of the PS - 19. Periodic (triangular), single shot, and manual state of the PS - 19. Periodic (triangular), single shot, and manual state of the PS - 19. Periodic (triangular), single shot, and manual state of the PS - 19. Periodic (triangular), single shot, and manual state of the PS - 19. Periodic (triangular), single shot, and manual state of the PS - 19. Periodic (triangular), single shot, and manual state of the PS - 19. Periodic (triangular), single shot, and manual state of the PS - 19. Periodic (triangular), single shot, and manual state of the PS - 19. Periodic (triangular), single shot, and manual state of the PS - 19. Periodic (triangular), single shot, and manual state of the PS - 19. Periodic (triangular), single shot, and manual state of the PS - 19. Periodic (triangular), single shot, and manual state of the PS - 19. Periodic (tri
Frequency setting at PS-19, digital keyboard, resolution
in steps by increment keys, smallest increment 1 Hz Error limits of frequency including ageing over 1 year
in steps by increment keys, smallest increment 1 Hz Error limits of frequency including ageing over 1 year
in steps by increment keys, smallest increment 1 Hz Error limits of frequency including ageing over 1 year
in steps by increment keys, smallest increment 1 Hz Error limits of frequency including ageing over 1 year
including ageing over 1 year
with Option Device BN 865/00.03
Slave tuning
Frequency setting at PSS-19 through Level Meters SPM-19, SPM-18 Level > $20.0\mathrm{dB}/\mathrm{dBm}$ $\pm 0.30\mathrm{dB}$ through Level Meters SPM-19, SPM-18 Level > $20.0\mathrm{dB}/\mathrm{dBm}$ $\pm 0.30\mathrm{dB}$ Level > $20.0\mathrm{dB}/\mathrm{dBm}$ $\pm 0.35\mathrm{dB}$ Output $Z_0 = 150\Omega$, 600Ω , $f = 200\mathrm{Hz}$ to $100\mathrm{kHz}$ $\pm 0.35\mathrm{dB}$ Output $Z_0 = 150\Omega$, 600Ω , $f = 200\mathrm{Hz}$ to $100\mathrm{kHz}$ $\pm 0.35\mathrm{dB}$ Output $Z_0 = 150\Omega$, 000Ω , $f = 200\mathrm{Hz}$ to $100\mathrm{kHz}$ $\pm 0.35\mathrm{dB}$ Individual errors (contained in total error) $Z_0 = 150\Omega$ at $Z_0 = 15$
through Level Meters SPM-19, SPM-18 Automatic frequency adjustments Auto-step Automatic incrementation of the frequency between two limits entered on the PS-19. Entry of increment, limits and speed of incrementation via keyboard. Sweep mode with PS-19, BN 870/02 and BN 870/22 with PSS-19 through SPM-19, BN 829/02, /03 or /22 Setting of upper and lower sweep limits or setting of centre frequency and sweep width Sweep periodic (triangular), single shot, and manual
Auto-step Automatic incrementation of the frequency between two limits entered on the PS-19. Entry of increment, limits and speed of incrementation via keyboard. Sweep mode with PS-19, BN 870/02 and BN 870/22 with PSS-19 through SPM-19, BN 829/02, /03 or /22 Setting of upper and lower sweep limits or setting of centre frequency and sweep width Sweep periodic (triangular), single shot, and manual Sweep $\frac{150 \Omega}{200 \Omega}$, $\frac{150 \Omega}{200 \Omega}$
Auto-step Automatic incrementation of the frequency between two limits entered on the PS-19. Entry of increment, limits and speed of incrementation via keyboard. Error limits of send level for $Z_{out} = Z_L = Z_o$ or $Z_{out} = 0$, $Z_L = Z_o$ at $f = 20$ kHz or 200 kHz (Tabulated values in dB) Sweep mode with PS-19, BN 870/02 and BN 870/22 with PSS-19 through SPM-19, BN 829/02, /03 or /22 Setting of upper and lower sweep limits or setting of centre frequency and sweep width Sweep periodic (triangular), single shot, and manual
Automatic incrementation of the frequency between two limits entered on the PS-19. Entry of increment, limits and speed of incrementation via keyboard. Error limits of send level for $Z_{out} = Z_L = Z_o$ or $Z_{out} = 0$, $Z_L = Z_o$ at $f = 20$ kHz or 200 kHz via keyboard. (Tabulated values in dB) Sweep mode with PS-19, BN 870/02 and BN 870/22 with PSS-19 through SPM-19, BN 829/02, /03 or /22 Setting of upper and lower sweep limits or setting of centre frequency and sweep width Sweep periodic (triangular), single shot, and manual
entered on the PS-19. Entry of increment, limits and speed of incrementation via keyboard. Sweep mode with PS-19, BN 870/02 and BN 870/22 with PSS-19 through SPM-19, BN 829/02, /03 or /22 Setting of upper and lower sweep limits or setting of centre frequency and sweep width Sweep periodic (triangular), single shot, and manual C_{constant}
with PS-19, BN 870/02 and BN 870/22 with PSS-19 through SPM-19, BN 829/02, /03 or /22
with PS-19, BN 870/02 and BN 870/22 with PSS-19 through SPM-19, BN 829/02, /03 or /22
with PS-19, BN 870/02 and BN 870/22 with PSS-19 through SPM-19, BN 829/02, /03 or /22
with PS-19, BN 870/02 and BN 870/22 with PSS-19 through SPM-19, BN 829/02, /03 or /22
with PSS-19 through SPM-19, BN 829/02, /03 or /22 $ Z_{\circ} = 75 \Omega $
Setting of upper and lower sweep limits or setting of centre frequency and sweep width Sweep periodic (triangular), single shot, and manual $Z_o = 124, 150 \Omega$ $Z_o = 124, 150 \Omega$ $Z_o = 150, 600 \Omega$
Sweep periodic (triangular), single shot, and manual $Z_o = 150,600 \Omega$ 20 kHz ± 0.18 ± 0.15
Sweep 1111 periodic (mangalan), single chot and manage
Sweep duration
for one half period
Error limits of the frequency response
Send level for $Z_{out} = Z_L = Z_o$ or $Z_{out} = 0$, $Z_L = Z_o$
Calibration switchable from voltage level (0 dB ≅ 0.775 V) (Tabulated values in dB)
to power level (0 dBm ≅ 1 mW in Z₀)
Out- referred
of absolute level, display in dBm, dB Z _o = 75 Ω 20 kHz ±0.25 ±0.10 ±0.08 ±0.10

1	referred tofat 8	30 200	Hz	1 6	iO 11	00 620	kHz 5	5 1	4 25 1	MHz
$Z_0 = 75 \Omega$	20 kHz	±0.25	±0.10		±0	.08	ı	±0	.10	
$Z_0 = 124,$ 150 Ω	200 kHz					±0.15		±0.20	_	
$Z_o = 150,$ 600Ω	20 kHz	±0.25 ²⁾	±0.15	±c	1.13	±0.25				

¹⁾ If $Z_{o}=50~\Omega,$ all level values given as reference are shifted by -7.5~dB or -5.7~dBm

or absolute power (voltage) level referred to a point

Fem. conn. (124 Ω) for WECO 372 A/379 A male conn. Fem. conn. (135 Ω) for WECO 241 A male conn. Fem. conn. (600 Ω) for WECO 310 A male conn.

1) BN 870/21 and /22: 135 Ω instead of 150 Ω

of zero relative level, display in dBm0, dB0 and the relative level, display in dBr

²⁾ for send level ≤0 dB/dBm

Spurious voltages

Harmonic ratio a_{k_2} , a_{k_3} for $Z_{out} = Z_1 = Z_0$ or $Z_{out} = 0$, $Z_1 = Z_0$	
Coaxial output, $Z_0 = 75 \Omega$, 800 Hz to 25 MHz	≧50 dB
Balanced outputs	> = 0 ID
$Z_{o} = 124 \Omega$, 150Ω ; 60kHz to 14MHz	
$Z_o = 150 \ \Omega, 600 \ \Omega; 800 \ Hz \ to \ 100 \ kHz $	≤50 ab
Suppression of discrete, nonharmonic	
spurious signals, in range 200 Hz to 25 MHz,	
signal-to-spurious signal ratio	≥60 dB
or spurious level ≦ – 120 dB (−1	
,	•

Memory

100 fixed frequencies (only PS-19) and 10 complete menus of instrument settings (PS-19, PSS-19) user programmable, entry and call-up via keyboard, erasure by writing over stored information.

General specifications

Power supply Rated ranges of use for a.c. line voltag	je, select	able
PS-19	96 to 140 227/237 V cy	0 V/193 to 261 V V, -12 to +10 % 47.5 to 63 Hz approx. 50 W
PSS-19	C	. арргох. 45 хА
Safety class as per IEC 348 and VDE 0	411	Class I
Tolerable ambient temperature Rated range of use	1	+5 to +40°C -40 to +70°C
Dimensions (in mm) and weight	alit	1050
Bench model	PS-19	PSS-19
Width with handles	. 199	477 155 432
19" conversion kit BN 70 Weight appro		BN 700/00.03 approx. 14 kg
German Post Office Certificate of Appl for PSS-19	1	No. 279 094 190 No. 279 094 195

Options

Higher Frequency Accuracy, BN 865/00.03	_
for PS-19, frequency error limits	$+1 \times 10^{-7}$

Interface < IEC 625 > Board, BN 853/05

for external control of all PS-19 instrument functions through an external computer.

Connection to the IEEE-Bus by adaptor plug S 834.

User-specific Memory, BN 870/00.01

for PS-19; storage of 100 fixed frequencies and 40 instrument settings, according to user's choice.

User-specific Memory, BN 871/00.01

for PSS-19; storage of 40 instrument settings according to user's choice.

Ordering Information

Level Generator PS-19

BN 870/..

BN 871/..

Standard version*	No sweep	BN 870/01
6.51	With sweep	BN 870/02
WECO connectors 1)	BN 870/21	
100	With sweep	BN 870/22

Send Section PSS-19 (for SPM-19 and SPM-18)

Standard version*	BN 871/01
WECO connectors 1)	BN 871/21

	WECO connectors 17	BN 871/21
4	Option (no extra charge) Generator output 50 Ω (instead of 75 Ω) ²⁾	BN 870/00.15
CC	Options (at extra cost) Higher Frequency Accuracy for PS-19 ²⁾	BN 865/00.03
	Interface < IEC 625 > Board for PS-19 with adaptor plug IEC 625/IEEE 488 (S 834) and connecting cable for IEEE 488 (K 420)	BN 853/05
	User-specific Memory for PS-19 ³⁾ User-specific Memory for PSS-19 ³⁾	BN 870/00.01 BN 871/00.01
	Accessories (at extra cost) Standard Frequency Adaptor SNZ-1 ⁴⁾ Connecting cable 24 pin, 50 cm long ⁵⁾	BN 956/00.07 K 366
	Front and back panel covers (1 set) SD-4 for PS-19 SD-3 for PSS-19	BN 700/00.24 BN 700/00.23

1) Impedances $Z_0=0,75,\,124,\,135,\,600\,\Omega$ 2) Factory fitted only. Please order with instrument.

 ³⁾ State chosen fixed frequencies (PS-19) and instrument settings when placing order (ask for Order Form No.5/785 a, b; or 5/785 a for the PSS-19).
 4) For specifications and extra ordering information see "Accessory" specification sheet.

⁵⁾ Cable is required for sweep and offset frequency measurements with PS-19/SPM-19.

Equipped with the 75 Ω basic connector Versacon® 9 and BNC adapter. For other adapter types, see "Specification Sheet Versacon® 9", and order chosen type when ordering instrument.